

November 25th, 2008

OUR FILE: 46405-3082

Submission No: 128287

CKC Laboratory Inc.  
22116 23<sup>rd</sup> Drive S.E.  
Bothell WA 98021  
USA

**Attention:** Steve Behm

Dear Sir/Madame:

The Bureau has received your application for the registration / renewal of a 3m alternate test site. Be advised that the information received was satisfactory to Industry Canada. The following number(s) is now associated to the site(s) for which registration / renewal was sought (**3082C-1**). Please reference the appropriate site number in the body of test reports containing measurements performed on the site. In addition, please be informed that the Bureau is now utilizing a **new site numbering scheme** in order to simplify the electronic filing process. Our goal is to reduce the number of secondary codes associated to one particular company. The following changes have been made to your records.

- Your primary code is: **3082**
- The following secondary code is now obsolete: **4653**
- The company number associated to the site(s) located at the above address is: **3082C**
- The table below is a summary of the changes made to the unique site registration number(s):


| <b>New Site Number</b> | <b>Obsolete Site Number</b> | <b>Description of Site</b> | <b>Expiry Date (YYYY-MM-DD)</b> |
|------------------------|-----------------------------|----------------------------|---------------------------------|
| 3082C-1                | 4653                        | 3m Chamber                 | 2010-11-25                      |

Furthermore, to obtain or renew a unique site number, the applicant shall demonstrate that the site has been accredited to ANSI C63.4-2003 or later. A scope of accreditation indicating the accreditation by a recognized accreditation body to ANSI C63.4-2003 shall be accepted. Please indicate in a letter the previous assigned site number if applicable and the type of site (example: 3 meter OATS or 3 meter chamber). If the test facility is not accredited to ANSI C63.4-2003 or later, the test facility shall submit test data demonstrating full compliance with the ANSI standard. The Bureau will evaluate the filing to determine if recognition shall be granted.

The frequency for re-validation of the test site and the information that is required to be filed or retained by the testing party shall comply with the requirements established by the accrediting organization. However, in all cases, test site re-validation shall occur on an interval not to exceed two years. There is no fee or form associated with an OATS filing. OATS submissions are encouraged to be submitted electronically to the Bureau using the following URL;  
[http://strategis.ic.gc.ca/epic/internet/inceb-bhst.nsf/en/h\\_tt00052e.html](http://strategis.ic.gc.ca/epic/internet/inceb-bhst.nsf/en/h_tt00052e.html).

If you have any questions, you may contact the Bureau by e-mail at [certification.bureau@ic.gc.ca](mailto:certification.bureau@ic.gc.ca)  
Please reference our file and submission number above for all correspondence.

Yours sincerely,



S. Proulx  
Wireless Laboratory Manager  
Certification and Engineering Bureau  
Industry Canada  
3701 Carling Ave., Building 94  
Ottawa, Ontario K2H 8S2  
Canada